



Australian Government

Department of Health and Aged Care
Therapeutic Goods Administration

CRITICAL QUARANTINE

LEVEL: Retail

CLASS: Class I

TGA REF: RC-2025-RN-00344-1

AGREEMENT DATE: 26/03/2025

PRODUCT: **Reynard Adhesive Removal Wipes, Reynard Alcohol Prep Pads and Reynard Alcohol/Chlorhexidine Swabsticks**

Multiple products and product codes

Batch numbers: All

ARTG's 309541, 309236 and 309237 (all cancelled)

SPONSOR: Reynard Health Supplies (Aus) Pty Ltd

REASON: The TGA has received reports of potential *Achromobacter* contamination. It has been determined a likely source of contamination is materials and products associated with the manufacturer Baoying County Fukang Medical Appliance Co Ltd.

While the investigation is ongoing into the potential contamination of bacteria *Achromobacter*, it has not been determined if the above products are affected. However, out of an abundance of caution, products from this manufacturer are to be quarantined until further notice.

**PROPOSED
CUSTOMER**

ACTIONS: Customers are to stop using these products and quarantine the affected stock on site until further notice.

This problem is still being investigated, and updates will be provided when available.

**SPONSOR
CONTACT**

INFORMATION: 1300 667 199 - Reynard Health Supplies

Please note that this notification is provided pursuant to subsections 61(3) and 61(7) of the Therapeutic Goods Act 1989.